

11-3-76 Purity Check on 11-1-1

Purpose: to make sure that this sample, which previously was dissolved in DMF, is still stable.

- inject 25 μ l

Obs. LC check valve in the proportionator has gone bad again, so can't run at high sens.


- it looked as though there was some early-eluting impurities - there also appeared to be some minor contam. in adduct region, but it was difficult to tell - **REDO THIS ANALYSIS WITH CLEAN COLUMN AND SAME SAMPLE - FIX LC VALVE!**

11-3-76 Harvest Crystals of Adduct.

It has been at least 24 hr, since the last prep run was completed. Crystals are present in all (4) fractions

- harvest crystals x cent in 15 ml tubes
- rinse w/ MeOH
- all crystals placed in (1) 15 ml tube

Column effluent fractions
containing adduct



↓
11-7-76 I noticed